

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031821\
 Data File : VD068657.D
 Acq On : 18 Mar 2021 16:34
 Operator : VA/SY
 Sample : M1686-02
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SW-SUMMER-STREET

Quant Time: Mar 19 01:06:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	217517	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	411573	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	416274	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	232261	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	127669	58.74	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.48%
35) Dibromofluoromethane	7.920	113	111769	45.08	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.16%
50) Toluene-d8	10.338	98	521488	52.81	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.62%
62) 4-Bromofluorobenzene	12.632	95	189060	59.40	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.80%
Target Compounds						
16) Acetone	4.156	43	4769	13.81	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031821\
Data File : VD068657.D
Acq On : 18 Mar 2021 16:34
Operator : VA/SY
Sample : M1686-02
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SW-SUMMER-STREET

Quant Time: Mar 19 01:06:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

