

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031821\
 Data File : VD068659.D
 Acq On : 18 Mar 2021 18:54
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 19 01:06:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	72	0.00
2 T	Dichlorodifluoromethane	0.485	0.472	2.7	69	0.00
3 P	Chloromethane	0.613	0.609	0.7	73	0.00
4 C	Vinyl Chloride	0.728	0.732	-0.5#	72	0.00
5 T	Bromomethane	0.518	0.505	2.5	74	0.00
6 T	Chloroethane	0.471	0.500	-6.2	78	0.00
7 T	Trichlorofluoromethane	0.831	0.904	-8.8	78	0.00
8 T	Diethyl Ether	0.247	0.289	-17.0	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.506	0.553	-9.3	79	0.00
10 T	Methyl Iodide	0.537	0.534	0.6	64	-0.01
11 T	Tert butyl alcohol	0.027	0.031	-14.8	93	0.00
12 CM	1,1-Dichloroethene	0.502	0.537	-7.0#	77	0.00
13 T	Acrolein	0.038	0.041	-7.9	75	0.00
14 T	Allyl chloride	0.648	0.683	-5.4	75	0.00
15 T	Acrylonitrile	0.104	0.113	-8.7	78	0.00
16 T	Acetone	0.079	0.079	0.0	70	0.00
17 T	Carbon Disulfide	1.645	1.587	3.5	69	0.00
18 T	Methyl Acetate	0.205	0.274	-33.7#	101	0.00
19 T	Methyl tert-butyl Ether	1.089	1.185	-8.8	79	0.00
20 T	Methylene Chloride	0.673	0.800	-18.9	95	0.00
21 T	trans-1,2-Dichloroethene	0.600	0.582	3.0	70	0.00
22 T	Diisopropyl ether	1.428	1.438	-0.7	72	0.00
23 T	Vinyl Acetate	0.741	0.767	-3.5	73	0.00
24 P	1,1-Dichloroethane	0.979	1.006	-2.8	74	-0.01
25 T	2-Butanone	0.114	0.121	-6.1	77	0.00
26 T	2,2-Dichloropropane	0.815	0.812	0.4	72	0.00
27 T	cis-1,2-Dichloroethene	0.658	0.673	-2.3	74	0.00
28 T	Bromochloromethane	0.382	0.400	-4.7	74	0.00
29 T	Tetrahydrofuran	0.074	0.079	-6.8	76	0.00
30 C	Chloroform	1.033	1.070	-3.6#	75	0.00
31 T	Cyclohexane	0.909	0.769	15.4	63	0.00
32 T	1,1,1-Trichloroethane	0.882	0.907	-2.8	74	0.00
33 S	1,2-Dichloroethane-d4	0.500	0.507	-1.4	73	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	70	0.00
35 S	Dibromofluoromethane	0.301	0.323	-7.3	76	0.00
36 T	1,1-Dichloropropene	0.475	0.469	1.3	69	0.00
37 T	Ethyl Acetate	0.158	0.158	0.0	74	0.00
38 T	Carbon Tetrachloride	0.447	0.463	-3.6	74	0.00
39 T	Methylcyclohexane	0.586	0.550	6.1	65	0.00
40 TM	Benzene	1.363	1.387	-1.8	72	0.00
41 T	Methacrylonitrile	0.092	0.091	1.1	75	0.00
42 TM	1,2-Dichloroethane	0.358	0.390	-8.9	76	0.00
43 T	Isopropyl Acetate	0.290	0.324	-11.7	78	0.00
44 TM	Trichloroethene	0.374	0.384	-2.7	73	0.00
45 C	1,2-Dichloropropane	0.326	0.354	-8.6#	75	0.00
46 T	Dibromomethane	0.177	0.196	-10.7	81	0.00
47 T	Bromodichloromethane	0.459	0.500	-8.9	77	0.00
48 T	Methyl methacrylate	0.154	0.152	1.3	70	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031821\
 Data File : VD068659.D
 Acq On : 18 Mar 2021 18:54
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 19 01:06:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	82	0.00
50 S	Toluene-d8	1.200	1.150	4.2	68	0.00
51 T	4-Methyl-2-Pentanone	0.152	0.171	-12.5	78	0.00
52 CM	Toluene	0.871	0.910	-4.5#	73	0.00
53 T	t-1,3-Dichloropropene	0.424	0.464	-9.4	78	0.00
54 T	cis-1,3-Dichloropropene	0.517	0.548	-6.0	75	0.00
55 T	1,1,2-Trichloroethane	0.245	0.275	-12.2	79	0.00
56 T	Ethyl methacrylate	0.282	0.311	-10.3	76	0.00
57 T	1,3-Dichloropropane	0.428	0.463	-8.2	77	0.00
58 T	2-Chloroethyl Vinyl ether	0.141	0.164	-16.3	81	0.00
59 T	2-Hexanone	0.101	0.116	-14.9	79	0.00
60 T	Dibromochloromethane	0.304	0.345	-13.5	79	0.00
61 T	1,2-Dibromoethane	0.240	0.270	-12.5	79	0.00
62 S	4-Bromofluorobenzene	0.387	0.459	-18.6	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	69	0.00
64 T	Tetrachloroethene	0.327	0.339	-3.7	74	0.00
65 PM	Chlorobenzene	1.018	1.091	-7.2	75	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.403	-13.8	80	0.00
67 C	Ethyl Benzene	1.818	1.938	-6.6#	74	0.00
68 T	m/p-Xylenes	0.690	0.747	-8.3	74	0.00
69 T	o-Xylene	0.636	0.717	-12.7	76	0.00
70 T	Styrene	1.087	1.238	-13.9	77	0.00
71 P	Bromoform	0.180	0.252	-40.0#	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.708	3.438	7.3	84	0.00
74 T	N-amyl acetate	0.632	0.632	0.0	89	0.00
75 P	1,1,2,2-Tetrachloroethane	0.638	0.654	-2.5	96	0.00
76 T	1,2,3-Trichloropropane	0.444	0.451	-1.6	109	0.00
77 T	Bromobenzene	0.846	0.801	5.3	88	0.00
78 T	n-propylbenzene	4.431	4.018	9.3	82	0.00
79 T	2-Chlorotoluene	2.459	2.321	5.6	86	0.00
80 T	1,3,5-Trimethylbenzene	3.039	2.996	1.4	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.190	0.192	-1.1	90	0.00
82 T	4-Chlorotoluene	2.588	2.612	-0.9	91	0.00
83 T	tert-Butylbenzene	2.588	2.568	0.8	89	0.00
84 T	1,2,4-Trimethylbenzene	3.044	3.126	-2.7	91	0.00
85 T	sec-Butylbenzene	3.656	3.650	0.2	90	0.00
86 T	p-Isopropyltoluene	3.318	3.358	-1.2	89	0.00
87 T	1,3-Dichlorobenzene	1.635	1.695	-3.7	93	0.00
88 T	1,4-Dichlorobenzene	1.619	1.668	-3.0	95	0.00
89 T	n-Butylbenzene	3.201	3.134	2.1	86	0.00
90 T	Hexachloroethane	0.660	0.523	20.8	69	0.00
91 T	1,2-Dichlorobenzene	1.402	1.493	-6.5	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.112	0.106	5.4	88	0.00
93 T	1,2,4-Trichlorobenzene	1.083	0.937	13.5	77	0.00
94 T	Hexachlorobutadiene	0.601	0.534	11.1	79	0.00
95 T	Naphthalene	1.961	1.812	7.6	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031821\
Data File : VD068659.D
Acq On : 18 Mar 2021 18:54
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Mar 19 01:06:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.937	0.845	9.8	82	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6