

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
 Data File : VD072286.D
 Acq On : 18 Mar 2022 13:25
 Operator : VA/SY
 Sample : N1956-03
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TPK-3

Quant Time: Mar 19 05:23:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	11449	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.867	114	18310	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	19553	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	9948	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	9416	73.797	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	147.600%
35) Dibromofluoromethane	7.909	113	7229	56.104	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.200%
50) Toluene-d8	10.332	98	22697	48.175	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.340%
62) 4-Bromofluorobenzene	12.620	95	8422	46.006	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
Data File : VD072286.D
Acq On : 18 Mar 2022 13:25
Operator : VA/SY
Sample : N1956-03
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TPK-3

Quant Time: Mar 19 05:23:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

