

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
 Data File : VD072289.D
 Acq On : 18 Mar 2022 14:50
 Operator : VA/SY
 Sample : N1966-10RE
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 G-1-10-18-19RE

Quant Time: Mar 19 05:24:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	8703	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	13586	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	14775	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	7348	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	6568	67.718	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.440%
35) Dibromofluoromethane	7.903	113	6295	65.843	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	131.680%
50) Toluene-d8	10.332	98	16303	46.635	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.280%
62) 4-Bromofluorobenzene	12.626	95	6181	45.505	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
Data File : VD072289.D
Acq On : 18 Mar 2022 14:50
Operator : VA/SY
Sample : N1966-10RE
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
G-1-10-18-19RE

Quant Time: Mar 19 05:24:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

