

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
 Data File : VD072292.D
 Acq On : 18 Mar 2022 16:16
 Operator : VA/SY
 Sample : N1995-07
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-08-4.5-5.0

Quant Time: Mar 19 05:25:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	189099	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	316322	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	383383	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	192213	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	126993	60.260	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.520%
35) Dibromofluoromethane	7.914	113	115616	51.939	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.880%
50) Toluene-d8	10.332	98	411873	50.603	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.200%
62) 4-Bromofluorobenzene	12.626	95	179114	56.636	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
Data File : VD072292.D
Acq On : 18 Mar 2022 16:16
Operator : VA/SY
Sample : N1995-07
Misc : 5.10G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-08-4.5-5.0

Quant Time: Mar 19 05:25:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

