

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
 Data File : VD072294.D
 Acq On : 18 Mar 2022 17:13
 Operator : VA/SY
 Sample : N1994-07
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-03-4.5-5.0RE

Quant Time: Mar 19 05:26:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	5754	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	9208	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	10785	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	4817	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	4650	72.515	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 145.020%	
35) Dibromofluoromethane	7.903	113	3882	59.910	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 119.820%	
50) Toluene-d8	10.332	98	11274	47.583	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 95.160%	
62) 4-Bromofluorobenzene	12.620	95	4700	51.053	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 102.100%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
Data File : VD072294.D
Acq On : 18 Mar 2022 17:13
Operator : VA/SY
Sample : N1994-07
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-03-4.5-5.0RE

Quant Time: Mar 19 05:26:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

