

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
 Data File : VD072296.D
 Acq On : 18 Mar 2022 18:11
 Operator : VA/SY
 Sample : N1994-13
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-05-4.5-5.0

Quant Time: Mar 19 05:26:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	874	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.861	114	916	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	717	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.57
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	289	29.671	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	59.340%
35) Dibromofluoromethane	7.920	113	447	69.346	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	138.700%
50) Toluene-d8	10.332	98	1233	52.313	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.620%
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
Data File : VD072296.D
Acq On : 18 Mar 2022 18:11
Operator : VA/SY
Sample : N1994-13
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-05-4.5-5.0

Quant Time: Mar 19 05:26:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

