

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
 Data File : VD072297.D
 Acq On : 18 Mar 2022 18:39
 Operator : VA/SY
 Sample : N1962-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 C-16-15-16

Quant Time: Mar 19 05:27:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	77632	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	129960	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	161779	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	83055	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	65805	76.061	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	152.120%
35) Dibromofluoromethane	7.908	113	53584	58.591	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	117.180%
50) Toluene-d8	10.338	98	163986	49.039	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.080%
62) 4-Bromofluorobenzene	12.620	95	75959	58.460	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	116.920%
Target Compounds						Qvalue
16) Acetone	4.167	43	11703	62.301	ug/l	92
40) Benzene	8.338	78	3898	1.065	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
Data File : VD072297.D
Acq On : 18 Mar 2022 18:39
Operator : VA/SY
Sample : N1962-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C-16-15-16

Quant Time: Mar 19 05:27:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

