

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
 Data File : VD072300.D
 Acq On : 18 Mar 2022 20:05
 Operator : VA/SY
 Sample : N1962-04
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 C-19-18-19

Quant Time: Mar 19 05:28:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	351945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	584568	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	639970	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	326848	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	169845	43.303	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	86.600%
35) Dibromofluoromethane	7.914	113	180306	43.831	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.660%
50) Toluene-d8	10.332	98	724386	48.159	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.320%
62) 4-Bromofluorobenzene	12.620	95	315854	54.043	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.080%
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	19160	22.499	ug/l #	86
17) Carbon Disulfide	4.403	76	21848	4.310	ug/l #	89
84) 1,2,4-Trimethylbenzene	13.255	105	32495	1.482	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
Data File : VD072300.D
Acq On : 18 Mar 2022 20:05
Operator : VA/SY
Sample : N1962-04
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C-19-18-19

Quant Time: Mar 19 05:28:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

