

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
 Data File : VD072320.D
 Acq On : 19 Mar 2022 06:29
 Operator : VA/SY
 Sample : N1970-13
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 G-4-3-6-7

Quant Time: Mar 19 08:10:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	24216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	39219	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	46624	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	23635	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	19345	71.682	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	143.360%
35) Dibromofluoromethane	7.914	113	15889	57.572	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.140%
50) Toluene-d8	10.338	98	46266	45.846	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.700%
62) 4-Bromofluorobenzene	12.626	95	21050	53.684	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.360%
Target Compounds						
					Qvalue	
95) Naphthalene	15.384	128	7359	8.371	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031822\
Data File : VD072320.D
Acq On : 19 Mar 2022 06:29
Operator : VA/SY
Sample : N1970-13
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
G-4-3-6-7

Quant Time: Mar 19 08:10:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

