

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031824\
 Data File : VD078694.D
 Acq On : 18 Mar 2024 14:02
 Operator : RP/MD
 Sample : P1746-19RE
 Misc : 5.07G/5ml/MSVOA_D/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-06-0-2.0RE

Quant Time: Mar 19 01:26:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 05:38:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	84401	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	146911	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	114701	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	31762	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	52106	65.876	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.760%
35) Dibromofluoromethane	7.804	113	54162	57.261	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.520%
50) Toluene-d8	10.269	98	172521	49.594	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.180%
62) 4-Bromofluorobenzene	12.569	95	32719	32.150	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	64.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031824\
Data File : VD078694.D
Acq On : 18 Mar 2024 14:02
Operator : RP/MD
Sample : P1746-19RE
Misc : 5.07G/5ml/MSVOA_D/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-06-0-2.0RE

Quant Time: Mar 19 01:26:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 05:38:03 2024
Response via : Initial Calibration

