

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031824\
 Data File : VD078703.D
 Acq On : 18 Mar 2024 18:06
 Operator : RP/MD
 Sample : P1786-01
 Misc : 4.77G/5ml/MSVOA_D/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-15

Quant Time: Mar 19 01:29:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 05:38:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	14699	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	32816	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	34428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	13501	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	17663	128.222	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 256.440%#	
35) Dibromofluoromethane	7.805	113	15297	72.400	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 144.800%	
50) Toluene-d8	10.269	98	38852	50.000	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.000%	
62) 4-Bromofluorobenzene	12.575	95	10861	47.777	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 95.560%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.811	84	2902	12.307	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031824\
Data File : VD078703.D
Acq On : 18 Mar 2024 18:06
Operator : RP/MD
Sample : P1786-01
Misc : 4.77G/5ml/MSVOA_D/SOIL/B
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-15

Quant Time: Mar 19 01:29:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 05:38:03 2024
Response via : Initial Calibration

