



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031824\
Data File : VD078688.D
Acq On : 18 Mar 2024 09:39
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
Title : SW846 8260
Last Update : Fri Mar 15 05:38:03 2024



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.0	6494	PASS
75	95	30	60	52.6	18981	PASS
95	95	100	100	100.0	36080	PASS
96	95	5	9	6.3	2270	PASS
173	174	0.00	2	1.6	560	PASS
174	95	50	100	97.9	35320	PASS
175	174	5	9	8.5	3007	PASS
176	174	95	101	96.5	34091	PASS
177	176	5	9	6.2	2105	PASS