

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD031919\
 Data File : VD061185.D
 Acq On : 19 Mar 2019 15:17
 Operator : VA/AP
 Sample : I.BLK
 Misc : 5.00µl/5ml/MSVOA D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleID :
 VD0320

Quant Time: Mar 20 01:59:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	708376	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.20	114	1137006	50.00	ug/l	-0.02
63) Chlorobenzene-d5	12.36	117	954963	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	423140	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.44	65	272957	46.29	ug/l	0.00
Spiked Amount			Recovery	=	92.58%	
35) Dibromofluoromethane	6.91	113	434635	51.51	ug/l	-0.02
Spiked Amount			Recovery	=	103.02%	
50) Toluene-d8	10.39	98	1049696	49.54	ug/l	0.00
Spiked Amount			Recovery	=	99.08%	
62) 4-Bromofluorobenzene	13.54	95	407520	49.56	ug/l	0.00
Spiked Amount			Recovery	=	99.12%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD031919\
Data File : VD061185.D
Acq On : 19 Mar 2019 15:17
Operator : VA/AP
Sample : I.BLK
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0320

Quant Time: Mar 20 01:59:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

