

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD031919\
 Data File : VD061195.D
 Acq On : 19 Mar 2019 19:53
 Operator : VA/AP
 Sample : K1935-09RE
 Misc : 3.08g/5ml/MSVOA D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-2(14-15)RE

Quant Time: Mar 20 02:21:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.03	168	326203	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.21	114	570255	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.35	117	396964	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	118531	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	167162	61.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.12%	
35) Dibromofluoromethane	6.92	113	252128	59.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.16%	
50) Toluene-d8	10.39	98	508898	47.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.76%	
62) 4-Bromofluorobenzene	13.54	95	132949	32.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.48%	
Target Compounds						
16) Acetone	3.22	43	17519	22.563	ug/l	93
17) Carbon Disulfide	3.38	76	63259	5.727	ug/l	96
20) Methylene Chloride	3.82	84	25975	2.948	ug/l #	90
39) Methylcyclohexane	8.86	83	7405	1.453	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD031919\
Data File : VD061195.D
Acq On : 19 Mar 2019 19:53
Operator : VA/AP
Sample : K1935-09RE
Misc : 3.08g/5ml/MSVOA D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-2(14-15)RE

Quant Time: Mar 20 02:21:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

