

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032122\
 Data File : VD072333.D
 Acq On : 21 Mar 2022 11:52
 Operator : VA/SY
 Sample : N1949-01
 Misc : 6.75G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CP-BH2-031422-2-3

Quant Time: Mar 22 02:10:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	397529	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	671116	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	724653	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	328003	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	210394	47.490	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.980%
35) Dibromofluoromethane	7.909	113	213988	45.311	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.620%
50) Toluene-d8	10.332	98	833963	48.294	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.580%
62) 4-Bromofluorobenzene	12.626	95	350699	52.267	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.540%
Target Compounds						
17) Carbon Disulfide	4.403	76	6228	2.209	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032122\
Data File : VD072333.D
Acq On : 21 Mar 2022 11:52
Operator : VA/SY
Sample : N1949-01
Misc : 6.75G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CP-BH2-031422-2-3

Quant Time: Mar 22 02:10:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

