

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032122\
 Data File : VD072342.D
 Acq On : 21 Mar 2022 16:07
 Operator : VA/SY
 Sample : N1970-09RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 G-3-12-21-22RE

Quant Time: Mar 22 02:12:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	62341	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	100364	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	116729	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	58864	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	39902	57.433	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.860%
35) Dibromofluoromethane	7.908	113	37261	52.758	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.520%
50) Toluene-d8	10.332	98	124484	48.203	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.400%
62) 4-Bromofluorobenzene	12.626	95	54599	54.412	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.820%
Target Compounds						
						Qvalue
16) Acetone	4.156	43	6972	46.220	ug/l	91
67) Ethyl Benzene	11.738	91	5017	1.085	ug/l #	86
95) Naphthalene	15.384	128	11839	5.407	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032122\
Data File : VD072342.D
Acq On : 21 Mar 2022 16:07
Operator : VA/SY
Sample : N1970-09RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
G-3-12-21-22RE

Quant Time: Mar 22 02:12:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

