

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032122\
 Data File : VD072345.D
 Acq On : 21 Mar 2022 17:33
 Operator : VA/SY
 Sample : N1962-01RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 C-16-15-16RE

Quant Time: Mar 22 03:15:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	21807	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	35481	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	44554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	22072	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	19279	79.329	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	158.660%
35) Dibromofluoromethane	7.909	113	14830	59.395	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.800%
50) Toluene-d8	10.332	98	45153	49.457	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.920%
62) 4-Bromofluorobenzene	12.626	95	19014	53.600	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.200%
Target Compounds						
16) Acetone	4.156	43	5871	111.265	ug/l #	80
25) 2-Butanone	7.209	43	1387	19.715	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032122\
Data File : VD072345.D
Acq On : 21 Mar 2022 17:33
Operator : VA/SY
Sample : N1962-01RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C-16-15-16RE

Quant Time: Mar 22 03:15:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

