

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032122\
 Data File : VD072347.D
 Acq On : 21 Mar 2022 18:29
 Operator : VA/SY
 Sample : N1995-01
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-06-4.5-5.0

Quant Time: Mar 22 02:13:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	65403	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	110055	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	132362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	68203	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	50789	69.681	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 139.360%	
35) Dibromofluoromethane	7.908	113	42830	55.303	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.600%	
50) Toluene-d8	10.332	98	134606	47.533	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 95.060%	
62) 4-Bromofluorobenzene	12.620	95	60892	55.340	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 110.680%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032122\
Data File : VD072347.D
Acq On : 21 Mar 2022 18:29
Operator : VA/SY
Sample : N1995-01
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-06-4.5-5.0

Quant Time: Mar 22 02:13:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

