

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032122\
 Data File : VD072351.D
 Acq On : 21 Mar 2022 20:23
 Operator : VA/SY
 Sample : N1968-10
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 G-2-11-23-24

Quant Time: Mar 22 02:14:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	584	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.855	114	958	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	569	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.349	152	178	50.000	ug/l	#-0.22
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	7.891	113	283	41.979	ug/l	-0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.960%
50) Toluene-d8	10.332	98	1158	46.977	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.960%
62) 4-Bromofluorobenzene	12.626	95	67	6.995	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	14.000%#
Target Compounds						
95) Naphthalene	15.373	128	160	24.167	ug/l	# 70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032122\
Data File : VD072351.D
Acq On : 21 Mar 2022 20:23
Operator : VA/SY
Sample : N1968-10
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
G-2-11-23-24

Quant Time: Mar 22 02:14:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

