

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032222\
 Data File : VD072366.D
 Acq On : 22 Mar 2022 17:01
 Operator : VA/SY
 Sample : N1995-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-06-4.5-5.0

Quant Time: Mar 23 01:13:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	332970	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	574464	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	645008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	321182	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	168270	45.346	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.700%
35) Dibromofluoromethane	7.908	113	183290	45.340	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.680%
50) Toluene-d8	10.332	98	734205	49.670	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.340%
62) 4-Bromofluorobenzene	12.626	95	307811	53.593	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032222\
Data File : VD072366.D
Acq On : 22 Mar 2022 17:01
Operator : VA/SY
Sample : N1995-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-06-4.5-5.0

Quant Time: Mar 23 01:13:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

