

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032222\
 Data File : VD072371.D
 Acq On : 22 Mar 2022 19:23
 Operator : VA/SY
 Sample : N2071-04
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PILE

Quant Time: Mar 23 01:15:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	394253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	668166	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	748896	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	356701	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	188227	42.840	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.680%
35) Dibromofluoromethane	7.914	113	200832	42.713	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.420%
50) Toluene-d8	10.332	98	860466	50.048	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.100%
62) 4-Bromofluorobenzene	12.620	95	379946	56.876	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032222\
Data File : VD072371.D
Acq On : 22 Mar 2022 19:23
Operator : VA/SY
Sample : N2071-04
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PILE

Quant Time: Mar 23 01:15:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

