

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032224\
 Data File : VD078736.D
 Acq On : 22 Mar 2024 12:10
 Operator : RP/MD
 Sample : VD0322SBL01
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0322SBL01

Quant Time: Mar 22 23:18:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 05:38:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	163468	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	307317	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	293679	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	131336	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	73310	47.854	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.700%
35) Dibromofluoromethane	7.810	113	93616	47.313	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.620%
50) Toluene-d8	10.269	98	375366	51.583	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.160%
62) 4-Bromofluorobenzene	12.575	95	102731	48.256	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.520%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032224\
Data File : VD078736.D
Acq On : 22 Mar 2024 12:10
Operator : RP/MD
Sample : VD0322SBL01
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0322SBL01

Quant Time: Mar 22 23:18:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 05:38:03 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.881 min Scan# 1069
Delta R.T. 0.006 min
Lab File: VD078736.D
Acq: 22 Mar 2024 12:10




Instrument :
MSVOA_D
ClientSampleId :
VD0322SBL01


Tgt Ion:168 Resp: 163468
Ion Ratio Lower Upper
168 100
99 51.2 38.8 58.2



#33
1,2-Dichloroethane-d4
Concen: 47.854 ug/l
RT: 8.234 min Scan# 1130
Delta R.T. -0.000 min
Lab File: VD078736.D
Acq: 22 Mar 2024 12:10

Tgt Ion: 65 Resp: 73310
Ion Ratio Lower Upper
65 100
67 54.4 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD078736.D

Acq: 22 Mar 2024 12:10

Instrument :

MSVOA_D

ClientSampleId :

VD0322SBL01





Tgt Ion:114 Resp: 307317

Ion Ratio Lower Upper

114 100

63 16.7 0.0 33.4

88 12.8 0.0 24.6



#35

Dibromofluoromethane

Concen: 47.313 ug/l

RT: 7.810 min Scan# 1058

Delta R.T. 0.006 min

Lab File: VD078736.D

Acq: 22 Mar 2024 12:10




Tgt Ion:113 Resp: 93616


Ion Ratio Lower Upper

113 100

111 104.7 81.9 122.9

192 19.1 15.4 23.0









#50
Toluene-d8
Concen: 51.583 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VD078736.D
Acq: 22 Mar 2024 12:10




Instrument :
MSVOA_D
ClientSampleId :
VD0322SBL01


Tgt Ion: 98 Resp: 375366
Ion Ratio Lower Upper
98 100
100 64.2 50.5 75.7



#62
4-Bromofluorobenzene
Concen: 48.256 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. 0.006 min
Lab File: VD078736.D
Acq: 22 Mar 2024 12:10

Tgt Ion: 95 Resp: 102731
Ion Ratio Lower Upper
95 100
174 98.2 0.0 195.2
176 91.4 0.0 188.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD078736.D
Acq: 22 Mar 2024 12:10

Instrument :
MSVOA_D
ClientSampleId :
VD0322SBL01

Tgt Ion:117 Resp: 293679
Ion Ratio Lower Upper
117 100
82 47.1 41.4 62.2
119 31.7 26.5 39.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. -0.000 min
Lab File: VD078736.D
Acq: 22 Mar 2024 12:10

Tgt Ion:152 Resp: 131336
Ion Ratio Lower Upper
152 100
115 54.9 27.1 81.2
150 156.7 0.0 350.2

