

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032319\
 Data File : VD061279.D
 Acq On : 23 Mar 2019 14:34
 Operator : VA/AP
 Sample : K2078-02
 Misc : 3.71g/5ml/MSVOA D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FK-SF-01-005-A

Quant Time: Mar 25 07:07:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	578727	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	8.23	114	1013818	50.00	ug/l	0.02
63) Chlorobenzene-d5	12.37	117	744884	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	14.52	152	279437	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.47	65	219753	45.62	ug/l	0.03
Spiked Amount	50.000		Recovery	=	91.24%	
35) Dibromofluoromethane	6.94	113	5295	0.70	ug/l	0.02
Spiked Amount	50.000		Recovery	=	1.40%	
50) Toluene-d8	10.42	98	751338	39.77	ug/l	0.03
Spiked Amount	50.000		Recovery	=	79.54%	
62) 4-Bromofluorobenzene	13.56	95	275865	37.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	75.26%	
Target Compounds						
					Qvalue	
16) Acetone	3.23	43	102532	74.432	ug/l	97
52) Toluene	10.52	92	21408	1.577	ug/l	98
68) m/p-Xylenes	12.67	106	16365	1.918	ug/l	99
69) o-Xylene	13.05	106	10563	1.259	ug/l	93
80) 1,3,5-Trimethylbenzene	14.24	105	70835	4.633	ug/l	97
84) 1,2,4-Trimethylbenzene	14.24	105	70835	4.633	ug/l	97
86) p-Isopropyltoluene	14.50	119	16882	1.011	ug/l	91
95) Naphthalene	16.12	128	61982	11.489	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032319\
Data File : VD061279.D
Acq On : 23 Mar 2019 14:34
Operator : VA/AP
Sample : K2078-02
Misc : 3.71g/5ml/MSVOA D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FK-SF-01-005-A

Quant Time: Mar 25 07:07:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

