

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032319\
 Data File : VD061292.D
 Acq On : 23 Mar 2019 20:33
 Operator : VA/AP
 Sample : K2056-04RE
 Misc : 7.94g/5ml/MSVOA D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CS-10RE

Quant Time: Mar 25 06:08:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	557155	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	8.23	114	998708	50.00	ug/l	0.01
63) Chlorobenzene-d5	12.37	117	728026	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	14.52	152	296921	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	238122	51.35	ug/l	0.01
Spiked Amount 50.000			Recovery	=	102.70%	
35) Dibromofluoromethane	6.93	113	378187	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
50) Toluene-d8	10.41	98	850582	45.70	ug/l	0.01
Spiked Amount 50.000			Recovery	=	91.40%	
62) 4-Bromofluorobenzene	13.55	95	277946	38.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.98%	
Target Compounds						
16) Acetone	3.23	43	9784	7.378	ug/l	94
20) Methylene Chloride	3.83	84	49934	3.964	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032319\
Data File : VD061292.D
Acq On : 23 Mar 2019 20:33
Operator : VA/AP
Sample : K2056-04RE
Misc : 7.94g/5ml/MSVOA D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CS-10RE

Quant Time: Mar 25 06:08:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

