


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD032319\
Data File : VD061274.D
Acq On : 23 Mar 2019 10:54
Operator : VA/AP
Sample : BFB
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Title : SW846 8260
Last Update : Fri Mar 08 02:35:39 2019



AutoFind: Scans 1233, 1234, 1235; Background Corrected with Scan 1229

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	17664	PASS
75	95	30	60	37.3	39138	PASS
95	95	100	100	100.0	104874	PASS
96	95	5	9	6.6	6906	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	82.7	86688	PASS
175	174	5	9	7.7	6646	PASS
176	174	95	101	97.4	84458	PASS
177	176	5	9	6.4	5381	PASS