

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032322\
 Data File : VD072389.D
 Acq On : 23 Mar 2022 17:41
 Operator : VA/SY
 Sample : N2063-09
 Misc : 5.58G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-8-(10-10.5)

Quant Time: Mar 24 00:53:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	334782	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	571957	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	647680	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	337476	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	170476	45.692	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.380%
35) Dibromofluoromethane	7.914	113	181450	45.082	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.160%
50) Toluene-d8	10.338	98	729521	49.570	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.140%
62) 4-Bromofluorobenzene	12.626	95	341383	59.699	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	119.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032322\
Data File : VD072389.D
Acq On : 23 Mar 2022 17:41
Operator : VA/SY
Sample : N2063-09
Misc : 5.58G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-8-(10-10.5)

Quant Time: Mar 24 00:53:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

