

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032322\
 Data File : VD072395.D
 Acq On : 23 Mar 2022 20:30
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Mar 24 00:55:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	479616	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	793135	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	899395	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	465501	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	207349	38.793	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	77.580%
35) Dibromofluoromethane	7.914	113	231651	41.504	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.000%
50) Toluene-d8	10.338	98	1048224	51.363	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.720%
62) 4-Bromofluorobenzene	12.620	95	475469	59.960	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	119.920%
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.314	83	36512	4.144	ug/l #	25
78) n-propylbenzene	12.808	91	47537	1.061	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032322\
Data File : VD072395.D
Acq On : 23 Mar 2022 20:30
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Mar 24 00:55:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

