


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD032420\  
 Data File : VD065517.D  
 Acq On : 24 Mar 2020 12:35  
 Operator : VA/SY  
 Sample : BFB  
 Misc : 5.00G/5.00ml/MSVOA D/SOIL  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA\_D\METHOD\82D030420S.M  
 Title : SW846 8260  
 Last Update : Thu Mar 05 01:22:07 2020



AutoFind: Scans 1822, 1823, 1824; Background Corrected with Scan 1813

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 50          | 95           | 15           | 40           | 17.5      | 55061   | PASS             |
| 75          | 95           | 30           | 60           | 48.0      | 151077  | PASS             |
| 95          | 95           | 100          | 100          | 100.0     | 314986  | PASS             |
| 96          | 95           | 5            | 9            | 6.8       | 21376   | PASS             |
| 173         | 174          | 0.00         | 2            | 1.2       | 3444    | PASS             |
| 174         | 95           | 50           | 100          | 92.8      | 292224  | PASS             |
| 175         | 174          | 5            | 9            | 7.5       | 22061   | PASS             |
| 176         | 174          | 95           | 101          | 98.5      | 287978  | PASS             |
| 177         | 176          | 5            | 9            | 6.7       | 19281   | PASS             |