

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032422\
 Data File : VD072405.D
 Acq On : 24 Mar 2022 14:00
 Operator : VA/SY
 Sample : N2094-03
 Misc : 5.28G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-2-(Q1)

Quant Time: Mar 25 02:54:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	379405	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	646752	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	684666	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	267597	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	200475	47.413	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.820%
35) Dibromofluoromethane	7.908	113	204936	45.029	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.060%
50) Toluene-d8	10.338	98	794126	47.719	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.440%
62) 4-Bromofluorobenzene	12.626	95	338574	52.361	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032422\
Data File : VD072405.D
Acq On : 24 Mar 2022 14:00
Operator : VA/SY
Sample : N2094-03
Misc : 5.28G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-2-(Q1)

Quant Time: Mar 25 02:54:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

