

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032422\
 Data File : VD072408.D
 Acq On : 24 Mar 2022 15:25
 Operator : VA/SY
 Sample : N2094-06
 Misc : 6.12G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-2-(Q4)

Quant Time: Mar 25 02:55:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	344916	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	600210	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	579724	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	132450	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	184641	48.035	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.060%
35) Dibromofluoromethane	7.908	113	194740	46.106	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.220%
50) Toluene-d8	10.337	98	697255	45.147	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	90.300%
62) 4-Bromofluorobenzene	12.626	95	252256	42.037	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032422\
Data File : VD072408.D
Acq On : 24 Mar 2022 15:25
Operator : VA/SY
Sample : N2094-06
Misc : 6.12G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-2-(Q4)

Quant Time: Mar 25 02:55:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

