

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032422\
 Data File : VD072413.D
 Acq On : 24 Mar 2022 17:47
 Operator : VA/SY
 Sample : N2070-04
 Misc : 5.93G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VNJ-212

Quant Time: Mar 25 02:57:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	361942	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	618688	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	720247	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	362919	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	188109	46.635	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.280%
35) Dibromofluoromethane	7.908	113	191605	44.009	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.020%
50) Toluene-d8	10.337	98	785077	49.315	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.640%
62) 4-Bromofluorobenzene	12.626	95	364552	58.935	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	117.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032422\
Data File : VD072413.D
Acq On : 24 Mar 2022 17:47
Operator : VA/SY
Sample : N2070-04
Misc : 5.93G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VNJ-212

Quant Time: Mar 25 02:57:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

