

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032519\
 Data File : VD061308.D
 Acq On : 25 Mar 2019 18:07
 Operator : VA/AP
 Sample : K2043-14
 Misc : 6.37g/5ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WP022SB433-190319-(23-25)

Quant Time: Mar 26 05:28:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032519S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 03:31:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	635762	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.23	114	1126705	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	820436	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	402241	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.47	65	216346	39.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.88%	
35) Dibromofluoromethane	6.94	113	354232	41.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.18%	
50) Toluene-d8	10.41	98	918498	43.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.88%	
62) 4-Bromofluorobenzene	13.55	95	309990	38.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.02%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.87	62	7512	1.584	ug/l	99
16) Acetone	3.24	43	33472	25.512	ug/l	98
20) Methylene Chloride	3.84	84	33272	4.468	ug/l	87
65) Chlorobenzene	12.40	112	241699	15.994	ug/l	100
87) 1,3-Dichlorobenzene	14.46	146	36823	2.843	ug/l	90
88) 1,4-Dichlorobenzene	14.54	146	81439	6.490	ug/l	99
91) 1,2-Dichlorobenzene	14.81	146	100122	9.687	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032519\
Data File : VD061308.D
Acq On : 25 Mar 2019 18:07
Operator : VA/AP
Sample : K2043-14
Misc : 6.37g/5ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WP022SB433-190319-(23-25)

Quant Time: Mar 26 05:28:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032519S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 03:31:57 2019
Response via : Initial Calibration

