

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032522\
 Data File : VD072436.D
 Acq On : 25 Mar 2022 18:35
 Operator : VA/SY
 Sample : N2095-03
 Misc : 5.08G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-03-5.0-6.0

Quant Time: Mar 26 00:15:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	13698	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	21529	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	27048	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	13688	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	9368	61.367	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 122.740%	
35) Dibromofluoromethane	7.914	113	8640	57.029	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 114.060%	
50) Toluene-d8	10.332	98	26857	48.481	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 96.960%	
62) 4-Bromofluorobenzene	12.620	95	12890	59.885	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 119.760%	
Target Compounds						
95) Naphthalene	15.379	128	10567	20.755	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032522\
Data File : VD072436.D
Acq On : 25 Mar 2022 18:35
Operator : VA/SY
Sample : N2095-03
Misc : 5.08G/5.0ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-03-5.0-6.0

Quant Time: Mar 26 00:15:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

