

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032621\
 Data File : VD068683.D
 Acq On : 26 Mar 2021 14:40
 Operator : VA/SY
 Sample : M1838-01
 Misc : 7.34G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PHC-BEAPT-001

Quant Time: Mar 27 00:43:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	253783	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	476663	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	407195	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	139688	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	158470	61.93	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.86%
35) Dibromofluoromethane	7.920	113	170611	57.53	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.06%
50) Toluene-d8	10.344	98	574074	50.21	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.42%
62) 4-Bromofluorobenzene	12.632	95	165207	44.56	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.12%
Target Compounds						
					Qvalue	
16) Acetone	4.162	43	20391	65.68	ug/l	94
17) Carbon Disulfide	4.409	76	21026	3.26	ug/l	95
25) 2-Butanone	7.209	43	6517	12.67	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032621\
Data File : VD068683.D
Acq On : 26 Mar 2021 14:40
Operator : VA/SY
Sample : M1838-01
Misc : 7.34G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PHC-BEAPT-001

Quant Time: Mar 27 00:43:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

