

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032621\
 Data File : VD068690.D
 Acq On : 26 Mar 2021 17:57
 Operator : VA/SY
 Sample : M1832-02
 Misc : 5.11G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 IG002

Quant Time: Mar 27 00:45:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	239504	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	468496	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	401124	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	130030	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	154263	63.88	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.76%
35) Dibromofluoromethane	7.920	113	167531	57.47	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.94%
50) Toluene-d8	10.338	98	563807	50.17	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.34%
62) 4-Bromofluorobenzene	12.632	95	157417	43.20	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.40%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032621\
Data File : VD068690.D
Acq On : 26 Mar 2021 17:57
Operator : VA/SY
Sample : M1832-02
Misc : 5.11G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
IG002

Quant Time: Mar 27 00:45:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

