

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032621\
 Data File : VD068691.D
 Acq On : 26 Mar 2021 18:25
 Operator : VA/SY
 Sample : M1832-03
 Misc : 6.65G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 IG003

Quant Time: Mar 27 00:45:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	271810	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	542152	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	504570	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	202533	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	160373	58.51	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.02%
35) Dibromofluoromethane	7.914	113	183119	54.29	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.58%
50) Toluene-d8	10.344	98	663037	50.98	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.96%
62) 4-Bromofluorobenzene	12.632	95	221742	52.59	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.18%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032621\
Data File : VD068691.D
Acq On : 26 Mar 2021 18:25
Operator : VA/SY
Sample : M1832-03
Misc : 6.65G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
IG003

Quant Time: Mar 27 00:45:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

