

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032621\
 Data File : VD068692.D
 Acq On : 26 Mar 2021 18:54
 Operator : VA/SY
 Sample : M1832-04
 Misc : 4.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 IG004

Quant Time: Mar 27 00:45:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	270461	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	534984	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	502134	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	200465	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	161425	59.19	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.38%
35) Dibromofluoromethane	7.920	113	181344	54.48	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.96%
50) Toluene-d8	10.344	98	665053	51.82	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.64%
62) 4-Bromofluorobenzene	12.632	95	211919	50.93	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.86%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032621\
Data File : VD068692.D
Acq On : 26 Mar 2021 18:54
Operator : VA/SY
Sample : M1832-04
Misc : 4.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
IG004

Quant Time: Mar 27 00:45:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

