

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD032719\
 Data File : VD061368.D
 Acq On : 28 Mar 2019 00:23
 Operator : VA/AP
 Sample : VD0327SBL02
 Misc : 5.00g/5ml/MSVOA_D/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0327SBL02

Quant Time: Mar 28 05:04:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 27 14:45:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.02	168	644704	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.19	114	1095886	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.35	117	934468	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.50	152	406309	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.43	65	315904	57.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.94%	
35) Dibromofluoromethane	6.89	113	515725	61.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.14%	
50) Toluene-d8	10.38	98	1168881	58.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.42%	
62) 4-Bromofluorobenzene	13.54	95	436620	56.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.38%	
Target Compounds						
20) Methylene Chloride	3.80	84	34642	3.846	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD032719\
Data File : VD061368.D
Acq On : 28 Mar 2019 00:23
Operator : VA/AP
Sample : VD0327SBL02
Misc : 5.00g/5ml/MSVOA_D/SOIL
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0327SBL02

Quant Time: Mar 28 05:04:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 27 14:45:32 2019
Response via : Initial Calibration

