

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD032720\
 Data File : VD065556.D
 Acq On : 27 Mar 2020 20:58
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 30 08:16:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032620S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 01:47:48 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	0.333	0.249	25.2#	85	0.00
3 P	Chloromethane	0.449	0.392	12.7	92	0.00
4 C	Vinyl Chloride	0.462	0.402	13.0#	87	0.00
5 T	Bromomethane	0.327	0.292	10.7	89	0.00
6 T	Chloroethane	0.299	0.275	8.0	90	0.00
7 T	Trichlorofluoromethane	0.689	0.613	11.0	87	0.00
8 T	Diethyl Ether	0.219	0.158	27.9#	69	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.455	0.332	27.0#	70	-0.01
10 T	Methyl Iodide	0.526	0.446	15.2	73	0.00
11 T	Tert butyl alcohol	0.023	0.024	-4.3	98	0.00
12 CM	1,1-Dichloroethene	0.432	0.312	27.8#	69	0.00
13 T	Acrolein	0.040	0.035	12.5	76	0.00
14 T	Allyl chloride	0.657	0.632	3.8	90	0.00
15 T	Acrylonitrile	0.090	0.093	-3.3	96	0.00
16 T	Acetone	0.079	0.059	25.3#	72	0.00
17 T	Carbon Disulfide	1.424	1.294	9.1	87	0.00
18 T	Methyl Acetate	0.204	0.203	0.5	93	0.00
19 T	Methyl tert-butyl Ether	0.861	0.900	-4.5	93	0.00
20 T	Methylene Chloride	0.609	0.498	18.2	95	0.00
21 T	trans-1,2-Dichloroethene	0.497	0.479	3.6	91	0.00
22 T	Diisopropyl ether	1.275	1.385	-8.6	97	0.00
23 T	Vinyl Acetate	0.694	0.752	-8.4	94	0.00
24 P	1,1-Dichloroethane	0.808	0.814	-0.7	95	0.00
25 T	2-Butanone	0.107	0.107	0.0	91	0.00
26 T	2,2-Dichloropropane	0.716	0.649	9.4	85	0.00
27 T	cis-1,2-Dichloroethene	0.515	0.524	-1.7	95	0.00
28 T	Bromochloromethane	0.336	0.359	-6.8	99	0.00
29 T	Tetrahydrofuran	0.069	0.072	-4.3	92	0.00
30 C	Chloroform	0.837	0.845	-1.0#	95	0.00
31 T	Cyclohexane	0.782	0.680	13.0	84	0.00
32 T	1,1,1-Trichloroethane	0.756	0.718	5.0	89	0.00
33 S	1,2-Dichloroethane-d4	0.416	0.484	-16.3	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.309	0.342	-10.7	105	0.00
36 T	1,1-Dichloropropene	0.457	0.418	8.5	87	0.00
37 T	Ethyl Acetate	0.175	0.177	-1.1	97	0.00
38 T	Carbon Tetrachloride	0.481	0.438	8.9	88	0.00
39 T	Methylcyclohexane	0.539	0.492	8.7	84	0.00
40 TM	Benzene	1.304	1.291	1.0	95	0.00
41 T	Methacrylonitrile	0.101	0.117	-15.8	116	0.00
42 TM	1,2-Dichloroethane	0.347	0.354	-2.0	99	0.00
43 T	Isopropyl Acetate	0.318	0.317	0.3	93	0.00
44 TM	Trichloroethene	0.378	0.354	6.3	91	0.00
45 C	1,2-Dichloropropane	0.317	0.322	-1.6#	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD032720\
 Data File : VD065556.D
 Acq On : 27 Mar 2020 20:58
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 30 08:16:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032620S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 01:47:48 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.166	0.171	-3.0	98	0.00
47 T	Bromodichloromethane	0.440	0.442	-0.5	98	0.00
48 T	Methyl methacrylate	0.151	0.152	-0.7	92	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	97	0.00
50 S	Toluene-d8	1.159	1.303	-12.4	102	0.00
51 T	4-Methyl-2-Pentanone	0.160	0.170	-6.3	94	0.00
52 CM	Toluene	0.826	0.836	-1.2#	95	0.00
53 T	t-1,3-Dichloropropene	0.403	0.412	-2.2	96	0.00
54 T	cis-1,3-Dichloropropene	0.493	0.497	-0.8	95	0.00
55 T	1,1,2-Trichloroethane	0.237	0.235	0.8	95	0.00
56 T	Ethyl methacrylate	0.257	0.273	-6.2	93	0.00
57 T	1,3-Dichloropropane	0.399	0.398	0.3	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.135	0.153	-13.3	99	0.00
59 T	2-Hexanone	0.109	0.112	-2.8	90	0.00
60 T	Dibromochloromethane	0.314	0.312	0.6	96	0.00
61 T	1,2-Dibromoethane	0.229	0.228	0.4	95	0.00
62 S	4-Bromofluorobenzene	0.368	0.422	-14.7	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.358	0.322	10.1	90	0.00
65 PM	Chlorobenzene	0.977	0.931	4.7	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.365	0.356	2.5	97	0.00
67 C	Ethyl Benzene	1.646	1.591	3.3#	91	0.00
68 T	m/p-Xylenes	0.646	0.624	3.4	92	0.00
69 T	o-Xylene	0.566	0.561	0.9	93	0.00
70 T	Styrene	1.000	1.021	-2.1	94	0.00
71 P	Bromoform	0.201	0.198	1.5	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.070	3.023	1.5	90	0.00
74 T	N-amyl acetate	0.606	0.616	-1.7	94	0.00
75 P	1,1,2,2-Tetrachloroethane	0.533	0.522	2.1	95	0.00
76 T	1,2,3-Trichloropropane	0.384	0.331	13.8	77	0.00
77 T	Bromobenzene	0.802	0.803	-0.1	96	0.00
78 T	n-propylbenzene	3.660	3.588	2.0	89	0.00
79 T	2-Chlorotoluene	2.088	2.029	2.8	91	0.00
80 T	1,3,5-Trimethylbenzene	2.545	2.552	-0.3	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.172	0.162	5.8	89	0.00
82 T	4-Chlorotoluene	2.219	2.156	2.8	92	0.00
83 T	tert-Butylbenzene	2.131	2.098	1.5	90	0.00
84 T	1,2,4-Trimethylbenzene	2.544	2.534	0.4	91	0.00
85 T	sec-Butylbenzene	3.050	2.905	4.8	88	0.00
86 T	p-Isopropyltoluene	2.811	2.755	2.0	90	0.00
87 T	1,3-Dichlorobenzene	1.544	1.478	4.3	93	0.00
88 T	1,4-Dichlorobenzene	1.543	1.442	6.5	92	0.00
89 T	n-Butylbenzene	2.552	2.412	5.5	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD032720\
Data File : VD065556.D
Acq On : 27 Mar 2020 20:58
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Mar 30 08:16:01 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032620S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 01:47:48 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.589	0.540	8.3	91	0.00
91 T	1,2-Dichlorobenzene	1.311	1.287	1.8	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.080	0.076	5.0	90	0.00
93 T	1,2,4-Trichlorobenzene	0.907	0.860	5.2	92	0.00
94 T	Hexachlorobutadiene	0.591	0.526	11.0	89	0.00
95 T	Naphthalene	1.360	1.364	-0.3	91	0.00
96 T	1,2,3-Trichlorobenzene	0.779	0.765	1.8	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6