

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032819\
 Data File : VD061393.D
 Acq On : 28 Mar 2019 16:18
 Operator : VA/AP
 Sample : K1950-02RE
 Misc : 3.67g/5ml/MSVOA D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FK-SF-03-005-ARE

Quant Time: Mar 28 17:01:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	564204	50.00	ug/l	0.03
34) 1,4-Difluorobenzene	8.22	114	987026	50.00	ug/l	0.03
63) Chlorobenzene-d5	12.36	117	663941	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	197103	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.45	65	261763	54.31	ug/l	0.02
Spiked Amount			Recovery	=	108.62%	
35) Dibromofluoromethane	6.92	113	180376	23.91	ug/l	0.02
Spiked Amount			Recovery	=	47.82%	
50) Toluene-d8	10.40	98	798966	44.16	ug/l	0.02
Spiked Amount			Recovery	=	88.32%	
62) 4-Bromofluorobenzene	13.55	95	251948	36.32	ug/l	0.00
Spiked Amount			Recovery	=	72.64%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.22	43	92778	74.883	ug/l	98
20) Methylene Chloride	3.82	84	44649	4.693	ug/l	94
40) Benzene	7.45	78	50064	2.316	ug/l	93
52) Toluene	10.50	92	44390	3.365	ug/l	92
68) m/p-Xylenes	12.66	106	23652	3.015	ug/l	100
69) o-Xylene	13.04	106	15828	2.079	ug/l	92
80) 1,3,5-Trimethylbenzene	13.93	105	49866	4.685	ug/l	95
84) 1,2,4-Trimethylbenzene	14.23	105	93575	8.121	ug/l	99
95) Naphthalene	16.12	128	101327	25.833	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032819\
Data File : VD061393.D
Acq On : 28 Mar 2019 16:18
Operator : VA/AP
Sample : K1950-02RE
Misc : 3.67g/5ml/MSVOA D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FK-SF-03-005-ARE

Quant Time: Mar 28 17:01:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

