

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032819\
 Data File : VD061394.D
 Acq On : 28 Mar 2019 16:46
 Operator : VA/AP
 Sample : K1950-04RE
 Misc : 3.92g/5ml/MSVOA D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FK-SF-03-005-BRE

Quant Time: Mar 28 17:13:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	559763	50.00	ug/l	0.03
34) 1,4-Difluorobenzene	8.21	114	971544	50.00	ug/l	0.02
63) Chlorobenzene-d5	12.36	117	725994	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	242140	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.45	65	256720	53.69	ug/l	0.02
Spiked Amount			Recovery	=	107.38%	
35) Dibromofluoromethane	6.92	113	124869	16.82	ug/l	0.02
Spiked Amount			Recovery	=	33.64%	
50) Toluene-d8	10.40	98	798614	44.85	ug/l	0.01
Spiked Amount			Recovery	=	89.70%	
62) 4-Bromofluorobenzene	13.55	95	266834	39.08	ug/l	0.00
Spiked Amount			Recovery	=	78.16%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.22	43	66709	54.269	ug/l #	87
20) Methylene Chloride	3.82	84	58693	6.867	ug/l	95
40) Benzene	7.45	78	42861	2.014	ug/l	97
52) Toluene	10.50	92	33017	2.543	ug/l	98
68) m/p-Xylenes	12.65	106	22130	2.580	ug/l	97
69) o-Xylene	13.05	106	12980	1.559	ug/l	98
80) 1,3,5-Trimethylbenzene	13.93	105	47231	3.612	ug/l	97
84) 1,2,4-Trimethylbenzene	14.23	105	83771	5.918	ug/l	97
95) Naphthalene	16.12	128	74624	15.487	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032819\
Data File : VD061394.D
Acq On : 28 Mar 2019 16:46
Operator : VA/AP
Sample : K1950-04RE
Misc : 3.92g/5ml/MSVOA D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FK-SF-03-005-BRE

Quant Time: Mar 28 17:13:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

