

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032819\
 Data File : VD061399.D
 Acq On : 28 Mar 2019 19:04
 Operator : VA/AP
 Sample : K2097-03RE
 Misc : 2.53g/5ml/MSVOA D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CS-22RE

Quant Time: Mar 29 01:42:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	465653	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	8.21	114	778371	50.00	ug/l	0.02
63) Chlorobenzene-d5	12.36	117	507391	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	138948	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	207433	52.15	ug/l	0.01
Spiked Amount 50.000			Recovery	=	104.30%	
35) Dibromofluoromethane	6.92	113	354343	59.57	ug/l	0.02
Spiked Amount 50.000			Recovery	=	119.14%	
50) Toluene-d8	10.41	98	629033	44.09	ug/l	0.02
Spiked Amount 50.000			Recovery	=	88.18%	
62) 4-Bromofluorobenzene	13.55	95	179521	32.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	65.62%	
Target Compounds						
16) Acetone	3.22	43	10559	10.326	ug/l	96
20) Methylene Chloride	3.82	84	51804	7.408	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032819\
Data File : VD061399.D
Acq On : 28 Mar 2019 19:04
Operator : VA/AP
Sample : K2097-03RE
Misc : 2.53g/5ml/MSVOA D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CS-22RE

Quant Time: Mar 29 01:42:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

