

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032819\
 Data File : VD061402.D
 Acq On : 28 Mar 2019 20:27
 Operator : VA/AP
 Sample : K2097-06RE
 Misc : 5.78g/5ml/MSVOA D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CS-6RE

Quant Time: Mar 29 01:46:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	30245	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	8.21	114	60801	50.00	ug/l	0.02
63) Chlorobenzene-d5	12.36	117	55275	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	23040	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	23764	91.98	ug/l	0.02
Spiked Amount 50.000			Recovery	=	183.96%	
35) Dibromofluoromethane	6.92	113	28785	61.95	ug/l	0.02
Spiked Amount 50.000			Recovery	=	123.90%	
50) Toluene-d8	10.40	98	54443	48.85	ug/l	0.02
Spiked Amount 50.000			Recovery	=	97.70%	
62) 4-Bromofluorobenzene	13.55	95	22633	52.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.92%	
Target Compounds						
20) Methylene Chloride	3.83	84	5484	13.331	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032819\
Data File : VD061402.D
Acq On : 28 Mar 2019 20:27
Operator : VA/AP
Sample : K2097-06RE
Misc : 5.78g/5ml/MSVOA D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CS-6RE

Quant Time: Mar 29 01:46:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

