

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032819\
 Data File : VD061420.D
 Acq On : 29 Mar 2019 7:07
 Operator : VA/AP
 Sample : K2107-05RE
 Misc : 2.58g/5ml/MSVOA D/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-3RE

Quant Time: Mar 29 08:26:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	22099	50.00	ug/l	0.04
34) 1,4-Difluorobenzene	8.23	114	49289	50.00	ug/l	0.04
63) Chlorobenzene-d5	12.36	117	49756	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	25077	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	21722	115.07	ug/l	0.03
Spiked Amount 50.000			Recovery	=	230.14%	
35) Dibromofluoromethane	6.93	113	24044	63.83	ug/l	0.03
Spiked Amount 50.000			Recovery	=	127.66%	
50) Toluene-d8	10.41	98	50175	55.54	ug/l	0.03
Spiked Amount 50.000			Recovery	=	111.08%	
62) 4-Bromofluorobenzene	13.55	95	25879	74.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	149.40%	
Target Compounds						
20) Methylene Chloride	3.84	84	8268	29.631	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032819\
Data File : VD061420.D
Acq On : 29 Mar 2019 7:07
Operator : VA/AP
Sample : K2107-05RE
Misc : 2.58g/5ml/MSVOA D/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-3RE

Quant Time: Mar 29 08:26:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

