

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032822\
 Data File : VD072449.D
 Acq On : 28 Mar 2022 17:26
 Operator : VA/SY
 Sample : N1977-06
 Misc : 4.78G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-14-20220315-15.0-16.0

Quant Time: Mar 29 01:30:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	3428	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.856	114	5452	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	5592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	1481	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	1409	36.882	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	73.760%
35) Dibromofluoromethane	7.903	113	2138	55.726	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.460%
50) Toluene-d8	10.332	98	6441	45.913	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.820%
62) 4-Bromofluorobenzene	12.626	95	1966	36.068	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	72.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032822\
Data File : VD072449.D
Acq On : 28 Mar 2022 17:26
Operator : VA/SY
Sample : N1977-06
Misc : 4.78G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-14-20220315-15.0-16.0

Quant Time: Mar 29 01:30:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

