

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032822\
 Data File : VD072453.D
 Acq On : 28 Mar 2022 19:19
 Operator : VA/SY
 Sample : N2103-03
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RBR-200055

Quant Time: Mar 29 01:31:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	416418	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	712046	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	818231	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	433003	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	203737	43.902	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.800%
35) Dibromofluoromethane	7.914	113	215651	43.038	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.080%
50) Toluene-d8	10.338	98	905402	49.417	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.840%
62) 4-Bromofluorobenzene	12.626	95	433562	60.902	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	121.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032822\
Data File : VD072453.D
Acq On : 28 Mar 2022 19:19
Operator : VA/SY
Sample : N2103-03
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RBR-200055

Quant Time: Mar 29 01:31:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

