

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032822\
 Data File : VD072454.D
 Acq On : 28 Mar 2022 19:47
 Operator : VA/SY
 Sample : N2103-05
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RBR-200031

Quant Time: Mar 29 01:31:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	107622	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	195349	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	240029	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	135737	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	75861	63.250	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.500%
35) Dibromofluoromethane	7.914	113	63952	46.521	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.040%
50) Toluene-d8	10.332	98	240830	47.912	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.820%
62) 4-Bromofluorobenzene	12.620	95	131522	67.340	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	134.680%
Target Compounds						
16) Acetone	4.156	43	5753	22.092	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032822\
Data File : VD072454.D
Acq On : 28 Mar 2022 19:47
Operator : VA/SY
Sample : N2103-05
Misc : 5.10G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RBR-200031

Quant Time: Mar 29 01:31:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

