

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032822\
 Data File : VD072444.D
 Acq On : 28 Mar 2022 11:01
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 29 01:28:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.462	0.366	20.8	67	0.00
3 P	Chloromethane	0.411	0.430	-4.6	88	0.00
4 C	Vinyl Chloride	0.397	0.445	-12.1#	88	0.00
5 T	Bromomethane	0.258	0.291	-12.8	97	0.00
6 T	Chloroethane	0.262	0.300	-14.5	93	0.00
7 T	Trichlorofluoromethane	0.896	0.822	8.3	77	0.00
8 T	Diethyl Ether	0.247	0.250	-1.2	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.571	0.519	9.1	77	0.00
10 T	Methyl Iodide	0.457	0.461	-0.9	79	0.00
11 T	Tert butyl alcohol	0.032	0.030	6.3	96	0.00
12 CM	1,1-Dichloroethene	0.460	0.447	2.8#	79	0.00
13 T	Acrolein	0.029	0.024	17.2	90	0.00
14 T	Allyl chloride	0.803	0.849	-5.7	88	0.00
15 T	Acrylonitrile	0.125	0.124	0.8	93	0.00
16 T	Acetone	0.121	0.126	-4.1	95	0.00
17 T	Carbon Disulfide	0.937	0.964	-2.9	72	0.00
18 T	Methyl Acetate	0.296	0.292	1.4	101	0.00
19 T	Methyl tert-butyl Ether	1.249	1.285	-2.9	92	0.00
20 T	Methylene Chloride	0.727	0.589	19.0	83	0.00
21 T	trans-1,2-Dichloroethene	0.508	0.516	-1.6	81	0.00
22 T	Diisopropyl ether	1.833	1.933	-5.5	93	0.00
23 T	Vinyl Acetate	0.935	1.029	-10.1	93	0.00
24 P	1,1-Dichloroethane	1.081	1.087	-0.6	87	0.00
25 T	2-Butanone	0.161	0.166	-3.1	96	0.00
26 T	2,2-Dichloropropane	0.975	0.959	1.6	85	0.00
27 T	cis-1,2-Dichloroethene	0.649	0.652	-0.5	86	0.00
28 T	Bromochloromethane	0.435	0.469	-7.8	94	0.00
29 T	Tetrahydrofuran	0.095	0.099	-4.2	95	0.00
30 C	Chloroform	1.151	1.138	1.1#	87	0.00
31 T	Cyclohexane	0.871	0.847	2.8	79	0.00
32 T	1,1,1-Trichloroethane	1.023	0.977	4.5	82	0.00
33 S	1,2-Dichloroethane-d4	0.557	0.538	3.4	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.352	0.332	5.7	95	0.00
36 T	1,1-Dichloropropene	0.494	0.483	2.2	81	0.00
37 T	Ethyl Acetate	0.216	0.211	2.3	90	0.00
38 T	Carbon Tetrachloride	0.561	0.512	8.7	79	0.00
39 T	Methylcyclohexane	0.555	0.548	1.3	78	0.00
40 TM	Benzene	1.408	1.403	0.4	85	0.00
41 T	Methacrylonitrile	0.126	0.137	-8.7	98	0.00
42 TM	1,2-Dichloroethane	0.427	0.437	-2.3	92	0.00
43 T	Isopropyl Acetate	0.420	0.423	-0.7	93	0.00
44 TM	Trichloroethene	0.395	0.378	4.3	81	0.00
45 C	1,2-Dichloropropane	0.386	0.388	-0.5#	91	0.00
46 T	Dibromomethane	0.190	0.188	1.1	89	0.00
47 T	Bromodichloromethane	0.554	0.538	2.9	88	0.00
48 T	Methyl methacrylate	0.200	0.201	-0.5	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032822\
 Data File : VD072444.D
 Acq On : 28 Mar 2022 11:01
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 29 01:28:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	92	0.00
50 S	Toluene-d8	1.287	1.182	8.2	89	0.00
51 T	4-Methyl-2-Pentanone	0.220	0.223	-1.4	94	0.00
52 CM	Toluene	0.927	0.916	1.2#	83	0.00
53 T	t-1,3-Dichloropropene	0.506	0.506	0.0	90	0.00
54 T	cis-1,3-Dichloropropene	0.585	0.588	-0.5	89	0.00
55 T	1,1,2-Trichloroethane	0.289	0.275	4.8	88	0.00
56 T	Ethyl methacrylate	0.339	0.340	-0.3	90	0.00
57 T	1,3-Dichloropropane	0.486	0.480	1.2	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.152	0.317	-108.6#	198#	0.00
59 T	2-Hexanone	0.152	0.161	-5.9	98	0.00
60 T	Dibromochloromethane	0.376	0.353	6.1	88	0.00
61 T	1,2-Dibromoethane	0.262	0.254	3.1	88	0.00
62 S	4-Bromofluorobenzene	0.500	0.460	8.0	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.352	0.327	7.1	79	0.00
65 PM	Chlorobenzene	1.108	1.062	4.2	84	0.00
66 T	1,1,1,2-Tetrachloroethane	0.438	0.416	5.0	85	0.00
67 C	Ethyl Benzene	1.980	1.952	1.4#	84	0.00
68 T	m/p-Xylenes	0.752	0.746	0.8	84	0.00
69 T	o-Xylene	0.719	0.711	1.1	84	0.00
70 T	Styrene	1.256	1.233	1.8	86	0.00
71 P	Bromoform	0.234	0.212	9.4	84	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.962	3.949	0.3	82	0.00
74 T	N-amyl acetate	0.851	0.907	-6.6	94	0.00
75 P	1,1,2,2-Tetrachloroethane	0.709	0.669	5.6	88	0.00
76 T	1,2,3-Trichloropropane	0.552	0.555	-0.5	91	0.00
77 T	Bromobenzene	0.892	0.865	3.0	84	0.00
78 T	n-propylbenzene	4.813	4.819	-0.1	83	0.00
79 T	2-Chlorotoluene	2.783	2.755	1.0	85	0.00
80 T	1,3,5-Trimethylbenzene	3.404	3.317	2.6	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.222	0.225	-1.4	92	0.00
82 T	4-Chlorotoluene	2.882	2.850	1.1	84	0.00
83 T	tert-Butylbenzene	2.984	2.896	2.9	82	0.00
84 T	1,2,4-Trimethylbenzene	3.354	3.296	1.7	84	0.00
85 T	sec-Butylbenzene	4.482	4.327	3.5	80	0.00
86 T	p-Isopropyltoluene	3.688	3.614	2.0	81	0.00
87 T	1,3-Dichlorobenzene	1.869	1.789	4.3	84	0.00
88 T	1,4-Dichlorobenzene	1.853	1.770	4.5	84	0.00
89 T	n-Butylbenzene	3.515	3.451	1.8	81	0.00
90 T	Hexachloroethane	0.723	0.682	5.7	82	0.00
91 T	1,2-Dichlorobenzene	1.630	1.531	6.1	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.113	0.108	4.4	87	0.00
93 T	1,2,4-Trichlorobenzene	1.079	0.995	7.8	82	0.00
94 T	Hexachlorobutadiene	0.628	0.550	12.4	76	0.00
95 T	Naphthalene	1.860	1.872	-0.6	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032822\
Data File : VD072444.D
Acq On : 28 Mar 2022 11:01
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Mar 29 01:28:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.934	0.857	8.2	82	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6