

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032824\
 Data File : VD078773.D
 Acq On : 28 Mar 2024 14:40
 Operator : RP/MD
 Sample : P1894-04
 Misc : 5.06G/5ml/MSVOA_D/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RB40858RT

Quant Time: Mar 29 01:34:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 05:38:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	138295	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	272347	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	241780	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	95448	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	66478	51.293	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.580%
35) Dibromofluoromethane	7.810	113	83657	47.709	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.420%
50) Toluene-d8	10.269	98	332481	51.557	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.120%
62) 4-Bromofluorobenzene	12.575	95	77742	41.207	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.420%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032824\
Data File : VD078773.D
Acq On : 28 Mar 2024 14:40
Operator : RP/MD
Sample : P1894-04
Misc : 5.06G/5ml/MSVOA_D/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RB40858RT

Quant Time: Mar 29 01:34:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 05:38:03 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.881 min Scan# 1069
Delta R.T. 0.006 min
Lab File: VD078773.D
Acq: 28 Mar 2024 14:40




Instrument :
MSVOA_D
ClientSampleId :
RB40858RT


Tgt Ion:168 Resp: 138295
Ion Ratio Lower Upper
168 100
99 50.2 38.8 58.2



#33
1,2-Dichloroethane-d4
Concen: 51.293 ug/l
RT: 8.234 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VD078773.D
Acq: 28 Mar 2024 14:40

Tgt Ion: 65 Resp: 66478
Ion Ratio Lower Upper
65 100
67 54.8 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD078773.D

Acq: 28 Mar 2024 14:40

Instrument :

MSVOA_D

ClientSampleId :

RB40858RT





Tgt Ion:114 Resp: 272347

Ion Ratio Lower Upper

114 100

63 16.9 0.0 33.4

88 12.9 0.0 24.6



#35

Dibromofluoromethane

Concen: 47.709 ug/l

RT: 7.810 min Scan# 1058

Delta R.T. 0.006 min

Lab File: VD078773.D

Acq: 28 Mar 2024 14:40




Tgt Ion:113 Resp: 83657


Ion Ratio Lower Upper

113 100

111 104.5 81.9 122.9

192 18.9 15.4 23.0









#50
Toluene-d8
Concen: 51.557 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD078773.D
Acq: 28 Mar 2024 14:40




Instrument :
MSVOA_D
ClientSampleId :
RB40858RT


Tgt Ion: 98 Resp: 332481
Ion Ratio Lower Upper
98 100
100 63.5 50.5 75.7



#62
4-Bromofluorobenzene
Concen: 41.207 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. 0.006 min
Lab File: VD078773.D
Acq: 28 Mar 2024 14:40

Tgt Ion: 95 Resp: 77742
Ion Ratio Lower Upper
95 100
174 98.1 0.0 195.2
176 91.7 0.0 188.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.581 min Scan# 1

Delta R.T. 0.000 min

Lab File: VD078773.D

Acq: 28 Mar 2024 14:40

Instrument :

MSVOA_D

ClientSampleId :

RB40858RT





Tgt Ion:117 Resp: 241780

Ion Ratio Lower Upper

117 100

82 49.4 41.4 62.2

119 31.6 26.5 39.7



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.516 min Scan# 2028

Delta R.T. 0.000 min

Lab File: VD078773.D

Acq: 28 Mar 2024 14:40

Tgt Ion:152 Resp: 95448

Ion Ratio Lower Upper

152 100

115 52.3 27.1 81.2

150 155.0 0.0 350.2

